
pupillometryData (PupillometryData)

b2boptic → items → item → pair → patient → pupillometryData



identification	
occurs	1

id (element of identification)	
type	string
occurs	1
restriction	length: 1..50 chars
description	measurement id from measurement device

timeStamp (element of identification)	
type	dateTime
occurs	1
description	measurement time from measurement device

counter (element of identification)	
type	int
occurs	0..1
restriction	min: 0
description	counter of ?

device (element of identification)	
occurs	0..1

deviceId (element of device)	
type	string

deviceId (element of device)	
occurs	1
restriction	length: 1..50 chars
description	id of measurement device

softwareVersion (element of device)	
type	string
occurs	1
restriction	length: 1..50 chars
description	software version of measurement device

pupilData	
occurs	0..4

side ***(attribute of pupilData)***	
type	Sides
use	required
description	eye side

condition ***(attribute of pupilData)***	
type	LightingContitions
use	required
description	lighting conditions

pupilDiameter (element of pupilData)	
type	float
unity	mm
occurs	1
description	diameter of pupil

pupilCenter (element of pupilData)	
occurs	1

x (element of pupilCenter)	
type	float
unity	mm
occurs	1
description	x coordinate of pupil center

y (element of pupilCenter)	
type	float
unity	mm
occurs	1
description	y coordinate of pupil center

application	
type	int
occurs	0..1
restriction	min: 0
description	

```
<xs:complexType name="PupillometryData">
  <xs:sequence>
    <xs:element name="identification">
      <xs:complexType>
```

```
      <xs:sequence>
        <xs:element name="id">
          <xs:simpleType>
            <xs:restriction base="xs:string">
              <xs:minLength value="1" />
              <xs:maxLength value="50" />
            </xs:restriction>
          </xs:simpleType>
        </xs:element>
        <xs:element name="timeStamp" type="xs:dateTime" />
        <xs:element minOccurs="0" name="counter">
          <xs:simpleType>
            <xs:restriction base="xs:int">
              <xs:minInclusive value="0" />
            </xs:restriction>
          </xs:simpleType>
        </xs:element>
        <xs:element minOccurs="0" name="device">
          <xs:complexType>
            <xs:sequence>
              <xs:element name="deviceId">
                <xs:simpleType>
                  <xs:restriction base="xs:string">
                    <xs:minLength value="1" />
                    <xs:maxLength value="50" />
                  </xs:restriction>
                </xs:simpleType>
              </xs:element>
              <xs:element name="softwareVersion">
                <xs:simpleType>
                  <xs:restriction base="xs:string">
                    <xs:minLength value="1" />
                    <xs:maxLength value="50" />
                  </xs:restriction>
                </xs:simpleType>
              </xs:element>
            </xs:sequence>
          </xs:complexType>
        </xs:element>
      </xs:sequence>
    </xs:complexType>
  </xs:element>
  <xs:element minOccurs="0" maxOccurs="4" name="pupilData">
    <xs:complexType>
      <xs:sequence>
        <xs:element name="pupilDiameter" type="xs:float" />
        <xs:element name="pupilCenter">
          <xs:complexType>
            <xs:sequence>
              <xs:element name="x" type="xs:float" />
              <xs:element name="y" type="xs:float" />
```

```
              </xs:sequence>
            </xs:complexType>
          </xs:element>
        </xs:sequence>
        <xs:attribute name="condition" type="LightingContitions"
use="required" />
        <xs:attribute name="side" type="Sides" use="required" />
      </xs:complexType>
    </xs:element>
    <xs:element minOccurs="0" name="application" type="xs:integer" />
  </xs:sequence>
</xs:complexType>
```

From:
http://wiki.b2boptic.com/ - **wiki.b2bOptic.com**

Permanent link:
http://wiki.b2boptic.com/en:testpages:t1

Last update: **2016/10/13 15:06**